

Korea Laboratory Accreditation Scheme

No.KRMPs-011

Branch site-1 : 87, Digital-ro 26-gil, Guro-gu, Seoul, Republic of Korea

1. Chemical Composition

109. Environmental pollutants

02. Air pollution

Type	(C)RM Code	Material (matrix etc)	Certified value or Range		Unit	Measurement Uncertainty (Confidence Level is about 95 %, $k = 2$)	Standard
CRM	KTL-CRM 109-01-1	Fine dust certified reference material for PAHs	Phenanthrene	100 ~ 5 000	mg/kg	60 %	NIER notice no. 2021-61
			Benzo [b]fluoranthene	1.0 ~ 100		60 %	air pollution test standard -ES 01803.1: 2021
CRM	KTL-CRM 109-01-2	Fine dust certified reference material for heavy metal analysis	Cr	1.0 ~ 100		60 %	NIER notice no. 2021-61 air pollution test standard -ES 01700: 2021
			Ni	1.0 ~ 100		60 %	
			Cd	10 ~ 500		60 %	
			Pb	1.0 ~ 100		60 %	

Korea Laboratory Accreditation Scheme

No.KRMPs-011

Branch site-2 : 723, Haean-ro, Sangnok-gu, Ansan-si, Gyeonggi-do, Republic of Korea

1. Chemical Composition

106. Inorganics, Rocks, Ores

06. Ceramics

Type	(C)RM Code	Material (matrix etc)	Certified value or Range		Unit	Measurement Uncertainty (Confidence Level is about 95 %, $k = 2$)	Standard
CRM	KTL-CRM 106-01-1	Insulating Ceramic certified reference material for heavy metal analysis	Pb	120 ~ 180	mg/kg	10 %	IEC 62321-5: 2013
			Cr	70 ~ 100		10 %	
			Cd	8 ~ 15		10 %	
CRM	KTL-CRM 106-01-2	Insulating Ceramic certified reference material for heavy metal analysis	Pb	750 ~ 1 100		10 %	
			Cr	700 ~ 1 000		10 %	
			Cd	80 ~ 110		10 %	
CRM	KTL-CRM 106-02-1	Barium Titanate certified reference material for heavy metal analysis	Pb	85 ~ 115		10 %	IEC 62321-5: 2013
			Cr	80 ~ 120		10 %	
			Cd	17 ~ 24		10 %	
CRM	KTL-CRM 106-02-2	Barium Titanate certified reference material for heavy metal analysis	Pb	850 ~ 1 100		10 %	
			Cr	850 ~ 1 200		10 %	
			Cd	80 ~ 110		10 %	

Korea Laboratory Accreditation Scheme

No.KRMPs-011

1. Chemical Composition

106. Inorganics, Rocks, Ores

06. Ceramics

Type	(C)RM Code	Material (matrix etc)	Certified value or Range		Unit	Measurement Uncertainty (Confidence Level is about 95 %, $k = 2$)	Standard
RM	KTL-RM 106-04-1	Zirconia reference material for chemical analysis	Al	900 ~ 1500	mg/kg	20 %	BS EN 725-12:2001
			Ti	30 ~ 70		20 %	
			Hf	10 000 ~ 20 000		10 %	
			Y	30 000 ~ 50 000		10 %	
RM	KTL-RM 106-04-2	Zirconia reference material for chemical analysis	Ti	20 ~ 60		20 %	BS EN 725-12:2001
			Hf	10 000 ~ 20 000		10 %	
			Y	90 000 ~ 110 000		10 %	
RM	KTL-RM 106-05-1	Yttrium oxide reference material for chemical analysis	Al	5 ~ 30		20 %	KTL L 137-2020A
			Fe	1 ~ 20		20 %	

07. Glasses

Type	(C)RM Code	Material (matrix etc)	Certified value or Range		Unit	Measurement Uncertainty (Confidence Level is about 95 %, $k = 2$)	Standard
CRM	KTL-CRM 106-03-1	Glasses certified reference material for heavy metal analysis	Pb	1 000 ~ 1 500	mg/kg	10 %	IEC 62321-5:2013
			Cr	100 ~ 150		10 %	
			Cd	50 ~ 100		10 %	
CRM	KTL-CRM 106-03-2	Glasses certified reference material for heavy metal analysis	Pb	200 ~ 300		10 %	
			Cr	15 ~ 25		10 %	
			Cd	5 ~ 15		10 %	

Korea Laboratory Accreditation Scheme

No.KRMPs-011

1. Chemical Composition

113. High molecular substance

02. High molecular substance additives

Type	(C)RM Code	Material (matrix etc)	Certified value or Range		Unit	Measurement Uncertainty (Confidence Level is about 95 %, $k = 2$)	Standard
CRM	KTL-CRM 113-01-1	ABS certified reference material for bromine-based flame retardant analysis	Decabromo diphenyl ether (BDE-209)	800 ~ 1200	mg/kg	10 %	IEC 6231-6:2015
CRM	KTL-CRM 113-01-2	ABS certified reference material for bromine-based flame retardant analysis	Decabromo diphenyl ether (BDE-209)	90 ~ 140		10 %	IEC 6231-6:2015
RM	KTL-RM 113-02-1	ABS certified reference material for chlorine-based flame retardant analysis	Tris (2-chloroethyl) phosphate (TCEP)	600 ~ 1000		10 %	KS M 1083:2019
RM	KTL-RM 113-02-2	ABS certified reference material for chlorine-based flame retardant analysis	Tris (2-chloroethyl) phosphate (TCEP)	80 ~ 120		10 %	KS M 1083:2019
RM	KTL-RM 113-03-1	High concentration polyethylene reference material for heavy metal analysis	Sn	400 ~ 600		10 %	IEC 62321-5:2013
			Sb	1 500 ~ 2 500			
RM	KTL-RM 113-03-2	Low concentration polyethylene reference material for heavy metal analysis	Sn	50 ~ 200		10 %	IEC 62321-5:2013
			Sb	300 ~ 700			

Korea Laboratory Accreditation Scheme

No.KRMPs-011

2. Physical properties

203. Thermodynamic properties

02. Freezing point & melting point cell

Type	(C)RM Code	Material (matrix etc)	Certified value or Range	Unit	Measurement Uncertainty (Confidence Level is about 95 %, $k = 2$)	Standard
CRM	KTL-CRM 203-01-1	Gallium certified reference material for non-contact temperature measurement	27.8 ~ 31.8	℃	0.4	ASTM E 1256:2017
CRM	KTL-CRM 203-01-2	Indium certified reference material for non-contact temperature measurement	154.6 ~ 159.6		0.4	ASTM E 1256:2017
CRM	KTL-CRM 203-01-3	Tin certified reference material for non-contact temperature measurement	229.9 ~ 233.9		0.4	ASTM E 1256:2017

Korea Laboratory Accreditation Scheme

No.KRMPs-011

2. Physical properties

206. Electrical and magnetic properties

03. Permittivity

Type	(C)RM Code	Material (matrix etc)	Certified value or Range		Unit	Measurement Uncertainty (Confidence Level is about 95 %, $k = 2$)	Standard
CRM	KIL-CRM 206-01-1	Ceramic/plastic certified reference material for Permittivity measurement	dielectric constant	(100 Hz ~ 10 MHz) 2 ~ 5	-	1.10 % (relative value)	ASTM D150-18 BS 763:1993
			loss tangent	(100 Hz ~ 10 MHz) 0.000 01 ~ 0.005		232×10^{-6} (absolute value)	
CRM	KIL-CRM 206-01-2		dielectric constant	(100 Hz ~ 10 MHz) 5 ~ 7		1.56 % (relative value)	
			loss tangent	(100 Hz ~ 10 MHz) 0.000 01 ~ 0.005		160×10^{-6} (absolute value)	
CRM	KIL-CRM 206-01-3		dielectric constant	(100 Hz ~ 10 MHz) 7 ~ 11		1.97 % (relative value)	
			loss tangent	(100 Hz ~ 10 MHz) 0.000 01 ~ 0.005		378×10^{-6} (absolute value)	
CRM	KIL-CRM 206-02-1		dielectric constant	(100 Hz ~ 10 MHz) 2 ~ 5		1.29 % (relative value)	
			loss tangent	(100 Hz ~ 10 MHz) 0.000 01 ~ 0.005		154×10^{-6} (absolute value)	
CRM	KIL-CRM 206-02-2		dielectric constant	(100 Hz ~ 10 MHz) 5 ~ 7		2.00 % (relative value)	
			loss tangent	(100 Hz ~ 10 MHz) 0.000 01 ~ 0.005		158×10^{-6} (absolute value)	
CRM	KIL-CRM 206-02-3		dielectric constant	(100 Hz ~ 10 MHz) 7 ~ 11		2.79 % (relative value)	
			loss tangent	(100 Hz ~ 10 MHz) 0.000 01 ~ 0.005		163×10^{-6} (absolute value)	

Korea Laboratory Accreditation Scheme

No.KRMPs-011

2. Physical properties

207. Precision measurement

07. Density

Type	(C)RM Code	Material (matrix etc)	Certified value or Range	Unit	Measurement Uncertainty (Confidence Level is about 95 %, $k = 2$)	Standard
CRM	KTL-CRM 207-01-1	Liquid certified reference material for density measurement	0.7	g/cm ³	0.003	KSMISO 12185:1996
CRM	KTL-CRM 207-01-2		0.8		0.003	
CRM	KTL-CRM 207-01-3		1.0		0.003	
CRM	KTL-CRM 207-01-4		1.2		0.003	
CRM	KTL-CRM 207-01-5		1.3		0.003	
CRM	KTL-CRM 207-01-6		1.6		0.003	
CRM	KTL-CRM 207-02-1	Water certified reference material for sugar content measurement	10	Sucrose in Water (%)	0.05	KSM 005:2017
CRM	KTL-CRM 207-02-2		12		0.05	
CRM	KTL-CRM 207-02-3		20		0.05	
CRM	KTL-CRM 207-02-4		30		0.05	
CRM	KTL-CRM 207-02-5		40		0.05	
CRM	KTL-CRM 207-02-6		50		0.05	
CRM	KTL-CRM 207-02-7		60		0.05	

Korea Laboratory Accreditation Scheme

No.KRMPs-011

2. Physical properties

207. Precision measurement

08. Liquid viscosity

Type	(C)RM Code	Material (matrix etc)	Certified value or Range						Unit	Measurement Uncertainty (Confidence Level is about 95 % $k = 2$)	Standard
			viscosity (mm ² /s)			viscosity (mPa.s)					
			20 °C	30 °C	40 °C	20 °C	30 °C	40 °C			
CRM	KTL-CRM 207-03-1	Silicone certified reference material for viscosity measurement	2.0	1.7	1.5	1.8	1.5	1.3	k- viscosity (mm ² /s) viscosity (mPa.s)	1.0 %	KSA 081:2016
CRM	KTL-CRM 207-03-2		10	9	7	9	8	7		1.0 %	
CRM	KTL-CRM 207-03-3		50	40	35	48	38	32		1.0 %	
CRM	KTL-CRM 207-03-4		100	80	66	97	78	62		1.0 %	
CRM	KTL-CRM 207-03-5		500	410	320	490	390	320		1.0 %	
CRM	KTL-CRM 207-03-6		1000	780	630	970	770	610		1.0 %	
CRM	KTL-CRM 207-03-7		5000	4000	3100	4800	3800	3000		1.0 %	
CRM	KTL-CRM 207-03-8		10 000	8200	6700	9700	7700	6200		1.0 %	
CRM	KTL-CRM 207-03-9		100 000	80 000	67 000	97 000	77 000	60 000		1.0 %	
CRM	KTL-CRM 207-03-10		300 000	246 000	206 000	291 000	241 000	191 000		1.0 %	

Korea Laboratory Accreditation Scheme

No.KRMPs-011

2. Physical properties

207. Precision measurement

09. Other

Type	(C)RM Code	Material (matrix etc)	Certified value or Range	Unit	Measurement Uncertainty (Confidence Level is about 95 %, $k = 2$)	Standard
CRM	KTL-CRM 207-04-1	Microscale certified reference material for microscope calibration	Pitch : 0.9 ~ 1.10	μm	0.04	ISO 11952:2019
			Pitch : 1.90 ~ 2.10		0.06	
			Pitch : 4.8 ~ 5.2		0.2	
			Pitch : 9.7 ~ 10.3		0.3	

Korea Laboratory Accreditation Scheme

No.KRMPs-011

3. Engineering characteristics

301. Particle characteristics

01. Particle size

Type	(C)RM Code	Material (matrix etc)	Certified value or Range	Unit	Measurement Uncertainty (Confidence Level is about 95 %, $k = 2$)	Standard
CRM	KTL-CRM 301-01-1	Silver nano-particle certified reference material for particle size analysis	15 ~ 25	nm	25 %	IEC 22412:2017
CRM	KTL-CRM 301-01-2		50 ~ 70		10 %	
CRM	KTL-CRM 301-01-3		95 ~ 125		10 %	

End.